UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/716,360	11/18/2003	Robert J. Thomas	1440.2032-001	6678
HAMILTON, BROOK, SMITH & REYNOLDS, P.C. 530 VIRGINIA ROAD			EXAMINER	
			PATEL, NIHIR B	
P.O. BOX 9133 CONCORD, MA 01742-9133			ART UNIT	PAPER NUMBER
			3772	
			MAIL DATE	DELIVERY MODE
			05/01/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)				
	10/716,360	THOMAS ET AL.				
Office Action Summary	Examiner	Art Unit				
	NIHIR PATEL	3772				
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address				
	VIC OFT TO EVEIDE AMANITUU	C) OD TUUDTY (OO) DAYO				
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DA - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period value of the period for reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tim vill apply and will expire SIX (6) MONTHS from cause the application to become ABANDONE	lely filed the mailing date of this communication. (35 U.S.C. § 133).				
Status						
1)⊠ Responsive to communication(s) filed on <u>04.10</u>).2008.					
	action is non-final.					
3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is						
closed in accordance with the practice under E	x parte Quayle, 1935 C.D. 11, 45	33 O.G. 213.				
Disposition of Claims						
4)⊠ Claim(s) <u>1-68</u> is/are pending in the application.						
4a) Of the above claim(s) <u>14-68</u> is/are withdrawn from consideration.						
5) Claim(s) is/are allowed.						
6)⊠ Claim(s) <u>1-13</u> is/are rejected.						
7) Claim(s) is/are objected to.						
8) Claim(s) are subject to restriction and/or	r election requirement.					
Application Papers						
9)☐ The specification is objected to by the Examine	r.					
10) ☐ The drawing(s) filed on is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11)☐ The oath or declaration is objected to by the Ex	aminer. Note the attached Office	Action or form PTO-152.				
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).						
a) ☐ All b) ☐ Some * c) ☐ None of:						
1. Certified copies of the priority documents have been received.						
2. Certified copies of the priority documents have been received in Application No						
3. Copies of the certified copies of the priority documents have been received in this National Stage						
application from the International Bureau (PCT Rule 17.2(a)).						
* See the attached detailed Office action for a list of the certified copies not received.						
Attachment(s)	л п	(DTO 440)				
Notice of References Cited (PTO-892) Notice of Draftsperson's Patent Drawing Review (PTO-948)	4)					
3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date 7/18.05; 9/16/04; 5/5/04; 11/18/03.	5) Notice of Informal P 6) Other:					
1 apor 110(0)/11/10/10 Date 1/10.00, 0/10/07, 0/0/07, 11/10/00.	5/ <u> </u>					



Application No.

Art Unit: 3772

DETAILED ACTION

Election/Restrictions

1. Applicant's election of **group II (claims 1-13)** in the reply filed on **April 10th**, **2008** is acknowledged. Because applicant did not distinctly and specifically point out the supposed errors in the restriction requirement, the election has been treated as an election without traverse (MPEP § 818.03(a)).

Claims **14-68** are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected species, there being no allowable generic or linking claim. Election was made **without** traverse in the reply filed on April 10th, 2008.

Claim Rejections - 35 USC § 102

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- 3. Claims 1, 2, 6 and 7 are rejected under 35 U.S.C. 102(b) as being anticipated by Raemer (US 5,320,093).
- 4. **As to claim 1,** Raemer teaches an apparatus that comprises a source of carbon dioxide (see col. 5 lines 5-30); an assembly 32 for combining pressurized air with the carbon dioxide resulting in a gas mix (see col. 5 lines 10-20); a patient centric ventilatory space module coupled to the assembly providing the resulting gas mix for inhalation by a given target, the inhalation of the gas mix effecting respiratory stability of the target (see col. 3 lines 50-60; col. 4 lines 10-20).

Art Unit: 3772

5. **As to claim 2,** Raemer teaches an apparatus wherein the assembly includes a positive airway pressure module for providing the pressurized air (see col. 4 lines 17-67).

- 6. **As to claim 6,** Raemer teaches an apparatus wherein at least one of the source, the assembly and the PCVSM is computer processor **50** controlled to modulate concentration of carbon dioxide in the gas mix (see col. 6 lines 66-68; col. 7 lines 1-20).
- 7. **As to claim 7,** Raemer teaches an apparatus wherein the computer processor modulates concentration of carbon dioxide in the gas mix as a function of any combination of sensed concentration of carbon dioxide in the PCVSM, sensed target state and detected system changes (see col. 7).

Claim Rejections - 35 USC § 103

- 8. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.
- 9. The factual inquiries set forth in *Graham* v. *John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:
 - 1. Determining the scope and contents of the prior art.
 - 2. Ascertaining the differences between the prior art and the claims at issue.
 - 3. Resolving the level of ordinary skill in the pertinent art.
 - 4. Considering objective evidence present in the application indicating obviousness or nonobviousness.
- 10. Claims **4**, **5**, **8**, **9**, **11**, **12** and **13** are rejected under 35 U.S.C. 103(a) as being unpatentable over Raemer (US 5,320,093).

Art Unit: 3772

11. **As to claims 8, 9 and 13,** Raemer substantially discloses a method steps of providing a substantially low concentration of carbon dioxide (see col. 5 lines 5-30); and combining pressurized air with the carbon dioxide to form a gas mix having a stabilizing effects on breathing (see col. 5 lines 10-20), the pressurized air enabling the carbon dioxide at low concentrations in the gas mix to have stabilizing effects on target respiratory systems (see col. 3 lines 50-60; col. 4 lines 10-20).

The method steps would have been obvious because they would have resulted from the use of the device of Raemer.

- 12. **As to claims 4, 5, 11 and 12,** Raemer substantially discloses the claimed invention; see rejection of claims 1 and 8 above, but does not disclose carbon dioxide in the gas mix that is less than 2%. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Raemer's invention by providing carbon dioxide in the gas mix that is less than 2% in order to provide cleanest gas possible, since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, **105 USPQ 233.**
- 13. Claims **3 and 10** are rejected under 35 U.S.C. 103(a) as being unpatentable over Raemer (US 5,320,093) in view of Pauley (US 5,975,078).
- 14. **As to claims 3 and 10,** Raemer substantially discloses the claimed invention; see rejection of claims 1 and 8 above, but does not disclose a PCVSM that includes an incubator, a tent, a facemask, and a nasal cannula. Pauley teaches an apparatus that does provide a PCVSM that includes an incubator, a tent, a facemask, and a nasal cannula (see figure 1; col. 3 lines 1-5). Therefore, it would have been obvious to one having ordinary skill in the art at the time the

Art Unit: 3772

invention was made to modify Raemer's invention by providing a PCVSM that includes an incubator, a tent, a facemask, and a nasal cannula as taught by Pauley in order to prevent leakage and inhale maximum amount of gas possible.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to NIHIR PATEL whose telephone number is (571)272-4803. The examiner can normally be reached on 7:30 to 4:30 every other Fridays off.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Patricia Bianco can be reached on (571) 272-4940. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Nihir Patel/ Examiner, Art Unit 3772

/Patricia Bianco/

Supervisory Patent Examiner, Art Unit 3772

Application/Control Number: 10/716,360

Page 6

Art Unit: 3772